



Ref. Certif. No.

DE 2-038793

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Genetic Sequencer

Name and address of the applicant

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122, USA

Name and address of the manufacturer

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122, USA

Name and address of the factory

Illumina, Inc.  
25861 Industrial Blvd.  
Hayward CA 94545, USA

Ratings and principal characteristics

AC 100-240V, 50/60Hz; 400W; Class I

Trademark (if any)

Illumina

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

MiSeqDx

Additional information (if necessary may also be reported on page 2)

Replaces CB Certificate DE 2-038381.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010+A1  
IEC 61010-2-081:2019  
IEC 61010-2-101:2018  
for national differences see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31181842 054

This CB Test Certificate is issued by the National Certification Body



TÜVRheinland®

TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2023-03-22

Signature:

Dipl.-Ing. M. Glagla