



Ref. Certif. No.

DE 2-035539-M2

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Next Generation Sequencer

Name and address of the applicant

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122, USA

Name and address of the manufacturer

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122, USA

Name and address of the factory

See additional page(s)

Ratings and principal characteristics

AC 200-240V, 50/60Hz; 15A (2500W); Class I; Laser Class 1

Trademark (if any)

Illumina

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

NovaSeq 6000Dx

Additional information (if necessary may also be reported on page 2)

Reissue of CB Certificate DE 2-035539 dated 18.04.2022 due to modification (alternate components). See associated test report for details

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010+A1  
IEC 61010-2-010:2019  
IEC 61010-2-081:2019  
IEC 61010-2-101:2018  
IEC 60825-1:2014  
for national differences see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31790423 029

This CB Test Certificate is issued by the National Certification Body



TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2023-08-02

Signature:

James Howell

1. Illumina, Inc.  
25861 Industrial Blvd.  
Hayward CA 94545  
USA
2. Venture Manufacturing Services Inc.  
1621 Barber Lane  
Milpitas CA 95035  
USA

**Additional information (if necessary)**

Report Ref. No. : 31790423 029



Date: 2023-08-02

Signature:

James Howell